



RADIATION
SOURCE SIDE

FIRE WALL SYSTEMS, UL LISTINGS:

1. UL PARTITION DESIGN TO BE U430 AT ONE HOUR AND 2 HOUR RATED WALLS.

GENERAL NOTES:

1. BOX LOCATIONS TO BE CENTERED BETWEEN THE STUDS.
2. THIS DETAIL SIMILAR IF REQUIRED AT CEILING PENETRATIONS.

SPECIFICATION NOTES:

1. THE LEAD SHIELDING VALUE TO BE THE SAME AS THE SURROUNDING WALL, PARTITION OR CEILING. (SEE PLANS OR PHYSICIST RADIATION SHIELDING REPORT FOR LEAD THICKNESS VALUES). CUSTOMER / INSTALLER MUST VERIFY LEAD SHIELDING EQUIVALENCY REQUIREMENT PRIOR TO ORDERING.
2. ALL RAY-BAR SHEET LEAD IS 99.9% PURE MEETING FEDERAL SPC QQ-L-201F GRADE C AND ASTM B749 TYPE L51121 AND ALL APPLICABLE NCRP REPORTS.
3. ALL RAY BAR GYPSUM BOARD MEETS CLASSIFICATION ASTM C 36.
4. MINIMUM HEIGHT TO BE 7'-0" A.F.F. PER NCRP, (U.O.N.)

BOX PENETRATION SHIELDING @ STEEL STUD PARTITION/CEILING

①
3" = 1'-0"

Terms of Use: Certain designs and concepts are the exclusive proprietary property and work product of Ray-Bar Engineering provided in good faith as a courtesy for the architectural design of projects specifying, incorporating and utilizing genuine Ray-Bar products and assemblies. Ray-Bar designs are not to be copied, altered, plagiarized or utilized in any competitive or detrimental manner against Ray-Bar as protected by law and/or copyrights. "All Rights Reserved"



JOB NAME:	
LOCATION:	
ROOM NUMBER:	
SHEET ____ OF ____:	

BOX PENETRATION SHIELD ALT @STL. STUD

RAY-BAR ENGINEERING CORP.

Toll Free: 1(800) 444-XRAY / 1(800) 444-9729
 Telephone: (626) 969-1818
 24 Hour Fax: 1(800) 333-XRAY / 1(800) 333-9729
 website: www.raybar.com

2012-05-10

SCALE
3" = 1'-0"

LBG-PS2